

CERTIFICATION

AOAC Research Institute Performance Tested MethodsSM

Certificate No. **102301**

The AOAC Research Institute hereby certifies the method known as:

Oculer Rapid 930 CS-Check[™] Commercial Sterility Testing Vials

manufactured by

Oculer Ltd Unit 2 Shannonside Business Park Birdhill, Co. Tipperary Ireland

This method has been evaluated in the AOAC Research Institute *Performance Tested Methods*SM Program and found to perform as stated in the applicability of the method. This certificate indicates an AOAC Research Institute Certification Mark License Agreement has been executed which authorizes the manufacturer to display the AOAC Research Institute *Performance Tested Methods*SM certification mark on the above-mentioned method for the period below. Renewal may be granted by the Expiration Date under the rules stated in the licensing agreement.

Fridly ASto

Bradley A. Stawick, Senior Director Signature for AOAC Research Institute

Issue Date Expiration Date November 26, 2023 December 31, 2024

2275 Research Blvd., Ste. 300, Rockville, Maryland, USA Telephone: +1-301-924-7077 Fax: +1-301-924-7089 Internet e-mail: aoacri@aoac.org * World Wide Web Site: http://www.aoac.org

AUTHORS Maria Cecilia Soria and Colin Fricker	SUBMITTING COMPANY Oculer Ltd Unit 2 Shannonside Business Park Birdhill, Co. Tipperary Ireland					
METHOD NAME Oculer Rapid 930 CS-Check [™] Commercial Sterility Testing Vials (50 vials)	CATALOG NUMBER OC-CSC-050					
INDEPENDENT LABORATORY Campden BRI Station Road Chipping Campden Gloucestershire, GL55 6LC United Kingdom						
APPLICABILITY OF METHOD						
Target organism – Aerobic mesophilic microorganisms.						
Matrixes – Ultra-high temperature (UHT) milk.						
Performance claims – The study data were unable to detect a statistical difference in results between the Oculer Rapid 930 CS-Check [™] method (hold for 5 days at 30°C) and the culture-based reference method indicated by European Directive 92/46 Annex C Chapter 1.A.4.b (hold for 15 days at 30°C) (2) for UHT milk.						
ORIGINAL CERTIFICATION DATE October 30, 2023	CERTIFICATION RENEWAL RECORD New approval 2023.					
METHOD MODIFICATION RECORD NONE	SUMMARY OF MODIFICATION NONE					
Under this AOAC <i>Performance Tested MethodsSM</i> License Number, 031601 this method is distributed by: NONE	Under this AOAC <i>Performance Tested Methodssm</i> License Number, 031601 this method is distributed as: NONE					

PRINCIPLE OF THE METHOD (1)

The Oculer Rapid 930 CS-Check™ system uses vials which contain a broth able to support the growth of aerobic, mesophilic microorganisms (bacteria, yeasts and molds) present in heat processed liquid products. As growth of the organisms occurs, oxygen is used from within the media vial and this depletion is automatically detected using a platinum porphyrin sensor. The sensor is excited by incident light and the lifetime of the emitted light is measured by the instrument. The lifetime of the emitted light from each individual vial is measured in real time by the Oculer Rapid 930 Series instrument. The time to growth detection in the Oculer Rapid 930 CS-Check correlates to the level of microorganisms present in the sample, with higher levels of contamination having a shorter detection time.

DISCUSSION OF THE VALIDATION STUDY (1)

Examination of the performance of different batches of Oculer Rapid 930 CS-Check, the results showed that the vials were consistently produced and remain stable over a year of storage at room temperature (18-25°C). Furthermore, small variations in sample volume, changes in sample temperature (2-8°C or 18-25°C) and different mixing techniques had no impact on test performance. Performance of the test on three separate instruments showed consistent performance. A panel of 51 (aerobic and facultative) reference cultures were selected for the inclusivity study, based on their link to hygiene and spoilage and 50% of these cultures were isolated from food.

The reference method used in the matrix study was the European Directive 92/46 Annex C, Chapter 1.A.4.b for UHT milk using ISO 4833-1:2013 for the plate count at 30°C. The methodology included 15 days of preincubation at 30°C and plating the samples on mPCA for colony counts after 72 h of incubation for the reference method. For the Oculer (candidate method) pre-incubation was for five days followed by up to 48 h incubation on the Oculer instrument. Using dPOD analysis, no statistical differences were detected between the Oculer Rapid 930 CS-Check methods and the reference method. These results emphasize the importance of the Oculer Rapid 930 CS-Check as a rapid method for commercial sterility tests. Use of the Oculer system results in a minimum of nine days time saving when compared to the reference method. This is largely achieved by the much greater volume of sample (after pre-incubation) analyzed when the Oculer system is used as compared to the reference method.

Oculer Rapid 930 CS-Check™ Commercial Sterility Testing Vials, AOAC Performance Tested Methods^{5M} Certification Number 102301

						Oculer Rapid Check	
No.	Genus	Species	CRA [♭] Number	Origin	Level inoculated (cfu/carton)	Time to Threshold	Result
1	Raoultella	terrigena	17343	Raw milk	9.6	00:00	+
2	Enterobacter	cloacae	1472	Dried milk	1.9	00:00	+
3	Klebsiella	oxytoca	8387	Water	2.1	00:00	+
4	Kluyvera	ascorbata	17126	Industrial	2.2	00:08	+
5	Escherichia	adecarboxylata	5501	Skim milk powder	10	00:00	+
6	Klebsiella	trevisanii	NCIMB ^c 8606	Ropy cream	2.4	27:01:00	+
7	Pantoea	agglomerans	17030	Pasteurized milk	1.8	00:08	+
8	Aeromonas	hydrophila	8388	Tin of milk with a fishy odor	2.5	03:50	+
9	Escherichia	coli	1476	Dried milk	1.6	00:12	+
10	Rahnella	aqualtilis	16911	Drinking water	1.4	00:00	+
11	Bacillus	coagulans	16586	Sterilized milk	3.4	08:22	+
12	Bacillus	subtilis	16579	Environmental	8.3	27:06:00	+
13	Bacillus	weihenstephanensis	16578	Pasteurized milk	9	03:57	+
14	Bacillus	polymyxa	7747	Unknown	2.8	00:31	+
15	Bacillus	cereus	7746	Milk or cream	8.5	10:59	+
16	Bacillus	pseudomycoides	16382	Soil in Sweden	4.6	29:26:00	+
17	Bacillus	pumilus	655	Chilled chicken in white wine sauce	4.1	02:16	+
18	Lysinibacillus	sphaericus	7746	Unknown	2.3	01:17	+
19	Bacillus	thuringiensis	NCIMB 9134	Flour moth	10	00:17	+
20	Bacillus	licheniformis	6335	Pesto sauce	8	07:31	+
21	Paenibacillus	macerans	16488	Unknown	3.8	06:25	+
22	Brevibacillus	brevis	7748	Unknown	3.8	03:09	+
23	Aneurinibacillus	aneurinolyticus	7751	Unknown	2.2	04:09	+
24	Paenibacillus	pabuli	16606	Barley	2.8	04:02	+
25	Brevibacillus	aigri	7749	Unknown	2.7	01:52	+

Table 4. Inclusivity List. Continued (1)

						Oculer Rapid Check	
No.	Genus	Species	CRA ^ь Number	Origin	Level inoculated (cfu/carton)	Time to Threshold	Result
26	Staphylococcus	carnosus	1123	Goat's milk	2.2	04:47	+
27	Listeria	ivanovii	16045	Soft cheese	1.8	01:42	+
28	Streptococcus	thermophilus	7675	Pasteurized milk	1	3:34:00	+
29	Lactobacillus	acidophilus	3910	Dairy product	1.8	00:04	+
30	Carnobacterium	divergens	8999	Brie	2.4	00:44	+
31	Staphylococcus	saprophyticus	3503	Distilled water environmental	8	02:48	+
32	Micrococcus	luteus	1513	Air sample	8.7	03:36	+
33	Enterococcus	faecalis	272	Dried milk powder	8.4	1:45:00	+
34	Staphylococcus	cohnii	409/3026	Skin	2.8	01:42	+
35	Staphylococcus	aureus	314	Slow cheese	1.5	00:12	+
36	Staphylococcus	epidermis	16030	Runway and can seam	9.7	0:04:00	+
37	Pediococcus	pentosaceus	100	Brine	3.4	05:53	+
38	Listeria	monocytogenes 1/2a	1100	Stilton	1.8	03:45	+
39	Listeria	innocua	16828	Cheese factory	8.5	3:06:00	+
40	Staphylococcus	hominis	16029	Unknown	9.3	02:15	+
41	Lactococcus	lactis	16659	Green ham	2.2	00:00	+
42	Micrococcus	roseus	7775	Water	1.2	00:00	+
43	Streptococcus	lactis	1511	Dried milk powder	2	00:00	+
44	Enterococcus	malodoratus	16860	Gouda cheese	8.6	00:00	+
45	Enterococcus	pseudoavium	16852	Cow udder - bovine mastitis	9.6	00:00	+
46	Aureobasidium	pullulans	16148	Soft drinks factory	3.2	36:05:00	+
47	Byssochlamys	fulva	16668	Pasteurized fruit juice	3.2	14:15	+
48	Candida	krussei	629	Yogurt base	4.2	03:12	+
49	Kluyveromyces	marxianus	6749	Dairy isolate	4.3	05:09	+
50	Torulaspora	delbruekeii	16154	Spoiled yogurt	9	1:16:00	+
51	Aneurinibacillus	aneurinolyticus	7758	human	10	00:27	+

^aResult: "+" = strain detected, "-" = strain not detected. ^bCampden BRI Culture Collection, Chipping Campden, UK. ^cNational Collection of Industrial, Food, and Marine Bacteria, Aberdeen, UK.

Table 5. Matrix study: Oculer Rapid 930 CS-Check presumptive results vs. confirmed results (1)											
	Contamination CS-Check presumptive results CS-Check confirmed results										
Matrix	Strain	level (cfu/g) ^a	N ^b	xc	POD _{CP} ^d	95% Cl ^f	х	PODcc ^e	95% Cl ^f	dPOD _{CP} ^g	95% CI
UHT whole	Bacillus	0	5	0	0.00	0.00, 0.43	0	0.00	0.00, 0.43	0	-0.47, 0.47
	subtilis	0.8	20	9	0.45	0.26, 0.66	8	0.40	0.22, 0.61	0.05	-0.11, 0.21
milk	(CRA ^h 16597)	4.1	5	5	1.00	0.57, 1.00	5	1.00	0.57, 1.00	0	-0.47, 0.47

^a = Inoculum level per g of sample matrix.

^b= Number of test potions.

^cx = Number of positive test portions.

^dPOD_{CP} = Candidate method presumptive results divided by the total number of trials.

 $^{\rm e}{\rm POD}_{\rm CC}$ = Candidate method confirmed results divided by the total number of trials.

^{195%} CI = If the confidence interval of a dPOD does not contain zero, then the difference is statistically significant at the 5% level.

^gdPOD_C = Difference between the candidate method presumptive and confirmed POD values.

^hCRA = Campden BRI Culture Collection, Chipping Campden, UK.

Table 6. Matrix Study Oculer Rapid 930 CS-Check results vs. Reference method results (1)

		Contamination		CS-Check method				Reference	e method		
Matrix	Strain	level (cfu/g) ^a	N ^b	xc	PODc ^d	95% CI ^f	х	POD _R ^e	95% CI	dPOD _{CR} ^g	95% CI
UHT	Bacillus	0	5	0	0.00	0.00, 0.43	0	0.0	0.0, 0.43	0.0	-0.43, 0.43
whole	subtilis	0.8	20	8	0.4	0.22, 0.61	11	0.55	0.34, 0.74	-0.15	-0.41, 0.15
milk	(CRA ^h 16597)	4.1	5	5	1.00	0.57, 1.00	5	1.0	0.57, 1.0	0.0	-0.43, 0.43

^a = Inoculum level per g of sample matrix.

^b= Number of test potions.

^cx = Number of positive test portions.

^dPOD_c = Candidate method presumptive results confirmed positive divided by the total number of trials.

 $^{e}POD_{R}$ = Reference method confirmed results divided by the total number of trials.

¹95% CI = If the confidence interval of a dPOD does not contain zero, then the difference is statistically significant at the 5% level.

^gdPOD_{CR} = POD difference between the candidate method confirmed and the Reference method.

^hCRA = Campden BRI Culture Collection, Chipping Campden, UK.

REFERENCES CITED

1. Soria, M.C. and Fricker, C., Validation of the Oculer Rapid 930 CS-Check[™] for the Determination of Commercial Sterility in Ultria-high Temperature Milk, AOAC *Performance Tested Methods*^{5M} certification number 102301.

2. European Directive 92/46 Annex C Chapter 1.A.4.B for Ultra-high temperature (UHT) milk